

Tracheal Heat Moisture Exchanger *Specifications*

Pepper Medical #HME01P



No	Test Item	Standard	Test request	Testing data	Result
1	Resistance	N/A	Resistance @ 30 LPM, resistance not exceed 0.3kPa	0.04kPa	Conformed
		N/A	Resistance @ 60 LPM, resistance not exceed 0.5kPa	0.2kPa	Conformed
		N/A	Resistance @ 90 LPM, resistance not exceed 0.7kPa	0.2kPa	Conformed
2	Humidity output	N/A	23.4mg @ VT 250ml Humidity output not exceed 15.0mg/L	1hour: 11.1mg/L 24hour: 11.7mg/L	Conformed
		N/A	23.4mg @ VT 500ml, Humidity output not exceed 20.0mg/L	1hour: 1.1mg/L 24hour: 3.7mg/L	Conformed